

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130291</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Sean Fielding</u>	ID# <u>5187</u> Date <u>8/20/09</u>
A Agency <u>Denali National Park</u>	Phone # <u>683-9528</u>
Instrument Location <u>Denali NP- HQ</u>	
Alco S/N <u>2301553</u> Target Value <u>.083</u> High Pressure <u>600</u>	
B Alco Test Value Average <u>.084</u>	<u>.083</u>
1 st Alco	2 nd Alco
Signature <u>[Signature]</u>	
(OVER)	


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130291

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130291

AUGUST 27, 2009

OPERATOR'S NAME:
FIELDING, JEROME
INSTRUMENT NUMBER: 5189
ANALYST'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:

D.L. #: 1874567
TEST STRIP: 01010
CHECK REPORT: 082709
TEST RESULT:
ALCO TARGET VALUE: .083
ALCO S/N: XZ01553

083 ADJUSTED FOR 27.90 in
ALCO TARGET .077 13184
BLANK TEST .000 13185
INTERNAL STANDARD VERIFIED 13186
ALCO TO 27.90 in .074 13186
BLANK TEST .000 13186
SUBJECT SAMPLE .083 13187
ALCO TO 27.90 in .083 13188
BLANK TEST .000 13189

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130291

AUGUST 27, 2009
TIME 13111

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
WEIGHTS
SAMPLE CHARGER: 49c
BREATH TUBE: 42c
BROOMER: 27.90 in
FLOW DETECTOR: OKAY

TEST STRIP: OKAY
DETECTOR: OKAY
FILTERS: OKAY
MAINT. STANDARD: OKAY
CALIBRATION: OKAY

PRIMER TEST
10/17/01 - 10/17/01 - 10/17/01 - 10/17/01
10/17/01 - 10/17/01 - 10/17/01 - 10/17/01
10/17/01 - 10/17/01 - 10/17/01 - 10/17/01